

## Air Ionizer Verification Record

Ionizer Verification Sequence Number: 08-056

### WORKING STANDARD USED

Asset/ISO #:	Manufacturer:	Model:	Serial No.	Calibration Date:	Calibration Due:	Calibration By:
25746	ION	775	7626	6/18/07	6/18/08	JPL

### AIR IONIZER INFORMATION

Asset/ISO #:	Manufacturer:	Model:	Serial No.	Verification Date:	Verification Due:	Verification By:
25355	SIMCO	Aerosol PC	33676700	5-2-08	11-1-08	JPL
Inspector:	Location:	Owner:	Fail: Y/N ?	Cleaned: Y/N ?	Adjusted: Y/N ?	Prior Sequence#
Phinh Do	306/21D		N	N	N	NA

### VERIFICATION DATA

HBM Sensitivity Level: 50 (from Table 1)

Fan controller setting: Medium (High, Low, NA)

Distance of ionizer from the charge plate: 24"

Ionizer Float Potential Tolerance  $\pm$  50 Vdc. (from Table 1)

Measured Float Potential values recorded below.

1	2	3	4	5	Comments:
0 Vdc.	0 Vdc.	0 Vdc.	0 Vdc.	0 Vdc.	

Ionizer Discharge Voltage Range:  $\pm$  1000 Vdc to  $\pm$  50 Vdc (from Table 1)

Ionizer Discharge Time Tolerance: 20 seconds. (from Table 1)

Measured Discharge Time in second(s) and recorded values below.

1 (+1000 to +Vdc)	2 (+1000 to +Vdc)	3 (+1000 to +Vdc)	4 (+1000 to +Vdc)	5 (+1000 to +Vdc)	Comments:
9.1 sec	9.2 sec	8.9 sec	8.8 sec	9.8 sec	
1 (-1000 to -Vdc)	2 (-1000 to -Vdc)	3 (-1000 to -Vdc)	4 (-1000 to -Vdc)	5 (-1000 to -Vdc)	Comments:
12.9 sec	12.9 sec	13.7 sec	13.9 sec	14.0 sec	

**Record** any corrective action required to restored ionizer operation (cleaning, adjustment, replacement, etc.)

If Ionizer was replaced, indicate below the identification of replacement.

Asset/ISO #: \_\_\_\_\_ Manufacturer: \_\_\_\_\_ Model: \_\_\_\_\_ Serial No.: \_\_\_\_\_

Sequence number for verification of replacement Ionizer: \_\_\_\_\_

**Record** inspection schedule and rational for that schedule.